

PALMSHIELD NITRILE POWDER FREE EXAMINATION GLOVES



Non-Sterile Nitrile Powder Free Examination Gloves

Intended use

To protect patient and user from cross contamination or infection during diagnostic and therapeutic procedures

Features

Material: Nitrile
Color: Blue
Cuff Finishing: Beaded
Surface Type: Textured
Sizes: X-Small, Small, Medium, Large, X-Large

Packaging:

100 Gloves per box
10 Boxes per case



Pricing is based on volume and delivery. Please send all inquiries and request for quotes to:

 **CLEAR DISTRIBUTORS**

161 East Main Street
Denville NJ 07834
Att: Dave Barrett
Phone: 973-222-9331
dbarrett@cleardistributorsusa.com

PRODUCT SPECIFICATION - Powder Free Nitrile Examination Gloves

Product Information : Nitrile Powder Free Sky Blue 4.0 gm,240 mm. min / 4.0 gm +/- 0.20 as per Medium				
Product Name	Non-sterile Nitrile Powder Free Examination Gloves			
Intended Use	To conduct diagnostic and therapeutic procedure to protect patient and user from cross contamination or infection.			
Product Conformance	ASTM D6319 Current Revision and Customer Requirements			
Material	Nitrile			
Cuff Finishing	Beaded			
Color	Sky Blue color			
Type of uses	Ambidextrous			
Surface Type	Textured			
Product Specification:				
Dimension	Size	Palm Width (mm)	Length (mm)	Weight (g.)
	X-Small	< 80	240min	2.8 – 3.2
	Small	80 + 10	240min	3.3 – 3.7
	Medium	95 + 10	240min	3.8 – 4.2
	Large	110 + 10	240min	4.3 – 4.7
	Extra Large	≥ 110	240min	4.8 – 5.2
Thickness	Single wall thickness (mm)			
		Specification		
	Cuff	N/A		
	Palm	0.08 min		
	Finger	0.10 min		
Physical Properties		Un-aged	Aged	
		Spec	Spec	
	Tensile Strength (MPa)	14 min	14min	
	Ultimate Elongation (%)	500min	400min	
	Force at Break (N)	6.0 min	6.0 min	
Protein content	N/A			
Residual Powder (Ref. ASTM D6124)	2.0 mg/glove maximum			
Quality Specification AND – Accept No Defects * n– Number of sample	Inspection Criteria		Sampling Plan/AQL	
	Dimension		S2 – AQL 4.0	
	Physical Properties		S2 – AQL 4.0	
	Critical Defects -(Embedded Hair/Insect/Foreign Matter)		G1-AND	
	Barrier Defects (Water Tight Test)		G1 – AQL 2.5 or customer required	
	Visual Inspection – Major		G1 – AQL 2.5	
	Visual Inspection – Minor		G1 – AQL 4.0	
	Residual Powder		*n = 5 (average)	
	Packaging Appearance		*n = 5 (AND)	
Glove Physical Counting		*n = 5 (AND)		
Packaging Configuration and Labeling				
Packing	Gloves 100 (+/-2) pcs per Inner Dispenser (by weight) 10 Dispensers per Outer Carton			
Labeling	The dispensers and outer cartons shall have the following information/prints: Product description and Brand including size and net quantity of gloves. Name and Country of Origin of Manufacturer or Distributor. Single Use Only or the Symbol for single use and Non-sterile			
Lot Numbering System	Per the customer's requirement or in house as per packing SOP. Manufacturing and/or expiry date where applicable.			

